



Republic of the Philippines  
Department of Agriculture  
**BUREAU OF SOILS AND WATER MANAGEMENT**  
SRDC Bldg., Elliptical Road corner Visayas Avenue,  
Diliman, Quezon City

**SUPPLEMENTAL BID BULLETIN**

**ADDENDUM NO. 2021-00-006-02**

This Addendum is issued to modify items in the Bid Documents. This shall form an integral part of the Bid Documents for the contract **PROCUREMENT OF SUPPLY AND DELIVERY OF VARIOUS TECHNICAL AND SCIENTIFIC EQUIPMENT FOR CY 2021** with an ABC of **ONE MILLION SEVEN HUNDRED TWENTY THOUSAND PESOS (PHP 1,720,000.00)**.

**REVISION IN SECTION VII. TECHNICAL SPECIFICATION:**

**AGREEMENTS**

**Incubator: to be used in incubating petri dishes for microbial determination (bacteria and fungi)**

**Lighting intensity: 0-25,000 lux**

**Humidity: 25-90%**

**Capacity: 400- 500 Liters**

**Temp. Range (w/ lighting): 10-650C**

**Uniformity with 4 shelves**

**Humidity stability: ±5-7%RH**

**Lighting intensity: 0-20000 lux**

For guidance and information of all concerned.

**ENGR. SAMUEL M. CONTRERAS**

Chairperson, Bids and Awards Committee

January 6, 2021



Republic of the Philippines  
Department of Agriculture  
**BUREAU OF SOILS AND WATER MANAGEMENT**  
SRDC Bldg., Elliptical Road corner Visayas Avenue,  
Diliman, Quezon City

## **SUPPLEMENTAL BID BULLETIN**

**ADDENDUM NO. 2021-00-006-02**

This Addendum is issued to modify items in the Bid Documents. This shall form an integral part of the Bid Documents for the contract **PROCUREMENT OF SUPPLY AND DELIVERY OF VARIOUS TECHNICAL AND SCIENTIFIC EQUIPMENT FOR CY 2021** with an ABC of **ONE MILLION SEVEN HUNDRED TWENTY THOUSAND PESOS (PHP 1,720,000.00)**.

### **REVISION IN SECTION VII. TECHNICAL SPECIFICATION:**

#### **AGREEMENTS**

**Incubator: to be used in incubating petri dishes for microbial determination (bacteria and fungi)**

**Lighting intensity: 0-25,000 lux**

**Humidity: 25-90%**

**Capacity: 400- 500 Liters**

**Temp. Range (w/ lighting): 10-650C**

**Uniformity with 4 shelves**

**Humidity stability: ±5-7%RH**

**Lighting intensity: 0-20000 lux**

For guidance and information of all concerned.

**ENGR. SAMUEL M. CONTRERAS**  
Chairperson, Bids and Awards Committee  
January 6, 2021